User Manual

HearingAssist DC-01 UV Dry Box

Thank you for purchasing the Hearing Assist DC-01 Ultraviolet Dry Box/Dehumidifier.

This box will lengthen the lifetime of your hearing aids. It removes moisture from your hearing aids and sterilizes them with UV Sterilization technology.


Read this manual before using the device.

Kit includes:

1) 1 UV Dry Box
2) 1 USB Cable
3) 1 Plug Adapter
4) 1 User manual

Step 1: Connect USB Cable with the dry box, then plug the other side into a powered USB port or connect to the provided plug adapter and plug into a socket. When correctly powered on, the LED screen lights up and shows "00".





Step 2: Press the single button inside the dry box to select the drying time. There are two modes: 3 hour (3H) and 6 hour (6H). We recommend using the 3H mode during the daytime and 6H mode overnight.



Step 3: After you select the mode, put hearing aid into the dry box and close the cover. The dry box will turn on and begin working as follows:

- The UV lamp lights up after the case is closed for 5 seconds, and a small indicator light between the 3 (or 6) and H in your screen will flash to note the UV lamp is on.
- UV sterilization will continue for 15 minutes, and then turn off, allowing the rest of the programmed period (3 or 6 hours) to be allocated to drying.
- When the drying is finished, the device will turn to standby mode and the process is completed.



Additional Notes

* You will see the hours countdown (from 6, to 5 etc.)
* When the screen shows "00" the process has finished
* There is no difference in functionality between the 6 hour and 3 hour programs, just the amount of time.
* You may clean your dry box with a soft dry cloth
* Working temperature of the dry box is about 45C/113F
* If you open the door to the device during the drying process , it will stop the process. You will have to reset it again to continue.
* We recommend using the dry box at least once a week to keep your hearing aid in optimal condition.